



Paris, the 10<sup>th</sup> January 2020

N. /Réf.: SB/FA – 9

**Objet: Accreditation of the SJ MILK SAMPLER automatic sampler from SCHWARTE Group for milk sampling for payment according to its composition and its hygienic and sanitary quality.**

Sir,

The *Commission Nationale d'Assistance et d'Harmonisation des Laboratoires* (C.N.A.H.), consulted on January 3, 2020, has examined the initial application for approval of the SJ MILK SAMPLER automatic sampler from SCHWARTE Group.

Considering the results of the phase I and II tests of the accreditation process of the SJ MILK SAMPLER automatic sampler, we have the honour to inform you of CNAH decision which validates the conclusions of the Experts Board *Préleveurs automatiques*.

**Accreditation of the SJ MILK SAMPLER automatic sampler according to the mounting configurations detailed below:**

MOUNTING CONFIGURATIONS						VALIDATION DATE
Pumping system	Actual pumping rate (cubic meters per hour)	Minimum volume to make a sample (liters)	Delay time (seconds)	Pipe length upstream of the sample (meters)	Pipe diameter upstream of the sample (millimeters)	
Impeller DN50-65	30	60	8 secondes min	12m max	40 – 65	03/01/2020
Ejector DN65	48 à 60	100	8 secondes min	12m max	50 – 65	03/01/2020
Water ring pump DN65	48 à 60	100	8 secondes min	12m max	50 – 65	03/01/2020

The Ejector DN80 and Water Ring pump DN80 configurations, having flow rates greater than the flow rates tested during the testing phases of the authorization process, cannot be authorized for use without prior testing.

The documents relating to this automatic sampler (user manual) must be sent to the CNIEL Laboratory Service each time it is updated. Any modification made to the automatic sampler must be the subject of an accreditation revision request which will be studied by the C.N.A.H.

The list of sampling systems that can be used for sampling milk intended for milk payment analyses will be updated and available on the CNIEL website shortly.

This accreditation is valid for a period of 5 years.

Your faithfully.



Jean-Claude PERRIN  
President of the C.N.A.H.

*Copy for information to the members of the C.N.A.H*